



VOYAGER™ RX

Coronary Dilatation Catheter

| Stock Number | UPN Number | Inflated Balloon Diameter (mm) | Crossing Profile (in) | Nominal Pressure (atm) | Rated Burst Pressure (atm) | Proximal-Distal Shaft Diameters (F) |
|--------------|----------------|--------------------------------|-----------------------|------------------------|----------------------------|-------------------------------------|
| 8 mm | | | | | | |
| 1011392-08 | 00802526411250 | 2.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011393-08 | 00802526411311 | 2.25 | .024 | 8 | 14 | 2.0/2.4 |
| 1011394-08 | 00802526411373 | 2.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011395-08 | 00802526411434 | 2.75 | .024 | 8 | 14 | 2.0/2.4 |
| 1011396-08 | 00802526411786 | 3.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011397-08 | 00802526411540 | 3.25 | .026 | 8 | 14 | 2.0/2.4 |
| 1011398-08 | 00802526411601 | 3.50 | .026 | 8 | 14 | 2.0/2.4 |
| 1011399-08 | 00802526411663 | 3.75 | .026 | 8 | 14 | 2.0/2.8 |
| 1011400-08 | 00802526411724 | 4.00 | .026 | 8 | 14 | 2.0/2.8 |
| 12 mm | | | | | | |
| 1011391-12 | 00802526411236 | 1.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011392-12 | 00802526411267 | 2.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011393-12 | 00802526411328 | 2.25 | .024 | 8 | 14 | 2.0/2.4 |
| 1011394-12 | 00802526411380 | 2.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011395-12 | 00802526411441 | 2.75 | .024 | 8 | 14 | 2.0/2.4 |
| 1011396-12 | 00802526411496 | 3.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011397-12 | 00802526411557 | 3.25 | .026 | 8 | 14 | 2.0/2.4 |
| 1011398-12 | 00802526411618 | 3.50 | .026 | 8 | 14 | 2.0/2.4 |
| 1011399-12 | 00802526411670 | 3.75 | .026 | 8 | 14 | 2.0/2.8 |
| 1011400-12 | 00802526411731 | 4.00 | .026 | 8 | 14 | 2.0/2.8 |
| 15 mm | | | | | | |
| 1011391-15 | 00802526411243 | 1.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011392-15 | 00802526411274 | 2.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011393-15 | 00802526411335 | 2.25 | .024 | 8 | 14 | 2.0/2.4 |
| 1011394-15 | 00802526411397 | 2.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011395-15 | 00802526411458 | 2.75 | .024 | 8 | 14 | 2.0/2.4 |
| 1011396-15 | 00802526411502 | 3.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011397-15 | 00802526411564 | 3.25 | .026 | 8 | 14 | 2.0/2.4 |
| 1011398-15 | 00802526411625 | 3.50 | .026 | 8 | 14 | 2.0/2.4 |
| 1011399-15 | 00802526411687 | 3.75 | .026 | 8 | 14 | 2.0/2.8 |
| 1011400-15 | 00802526411748 | 4.00 | .026 | 8 | 14 | 2.0/2.8 |
| 20 mm | | | | | | |
| 1011392-20 | 00802526411281 | 2.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011393-20 | 00802526411342 | 2.25 | .024 | 8 | 14 | 2.0/2.4 |
| 1011394-20 | 00802526411403 | 2.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011395-20 | 00802526411465 | 2.75 | .024 | 8 | 14 | 2.0/2.4 |
| 1011396-20 | 00802526411519 | 3.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011397-20 | 00802526411571 | 3.25 | .026 | 8 | 14 | 2.0/2.4 |
| 1011398-20 | 00802526411632 | 3.50 | .026 | 8 | 14 | 2.0/2.4 |
| 1011399-20 | 00802526411694 | 3.75 | .026 | 8 | 14 | 2.0/2.8 |
| 1011400-20 | 00802526411755 | 4.00 | .026 | 8 | 14 | 2.0/2.8 |

| Stock Number | UPN Number | Inflated Balloon Diameter (mm) | Crossing Profile (in) | Nominal Pressure (atm) | Rated Burst Pressure (atm) | Proximal-Distal Shaft Diameters (F) |
|---|----------------|--------------------------------|-----------------------|----------------------------|----------------------------|-------------------------------------|
| 30 mm | | | | | | |
| 1011392-30 | 00802526411304 | 2.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011393-30 | 00802526411366 | 2.25 | .024 | 8 | 14 | 2.0/2.4 |
| 1011394-30 | 00802526411427 | 2.50 | .024 | 8 | 14 | 2.0/2.4 |
| 1011395-30 | 00802526411489 | 2.75 | .024 | 8 | 14 | 2.0/2.4 |
| 1011396-30 | 00802526411533 | 3.00 | .024 | 8 | 14 | 2.0/2.4 |
| 1011397-30 | 00802526411595 | 3.25 | .026 | 8 | 14 | 2.0/2.4 |
| 1011398-30 | 00802526411656 | 3.50 | .026 | 8 | 14 | 2.0/2.4 |
| 1011399-30 | 00802526411717 | 3.75 | .026 | 8 | 14 | 2.0/2.8 |
| 1011400-30 | 00802526411779 | 4.00 | .026 | 8 | 14 | 2.0/2.8 |
| Usable Catheter Working Length = 143 cm | | | | Maximum Guide Wire = .014" | | |

*Lee, DP et al. Optimizing Clinical Outcomes with Intracoronary Stenting: The Importance of Minimizing Vessel Injury. Stent 200 0. Vol 3 No 1.